

Czeladź, dn. 02.01.2013

**DECLARATION OF CONFORMITY**

**Manufacturer: Reha – Bed Sp. z o.o.**  
**Ul. Spacerowa 1**  
**41-253 Czeladź**

**Declare medical devices:**

**Nursing care bed type TAURUS**



It is hereby declared that the medical equipment listed above, marked CE is medical device of First Class, rule 12, in accordance with Directive 93/42/EWG<sup>1</sup> (adopted into a Polish law in Medical Equipment Act of 20 May 2010)

In Reha-Bed, there is implemented the Quality Management System conformable with the requirements of the standard PN-EN ISO 9001:2009.


The procedure of conformity definite fundamental claims included in Annex 1 and in accordance with Annex 7 of Minister of Health Order of 12 January 2011


Medical equipment is in conformity with the following European standards:

- PN-EN ISO 14971:2011
- PN-EN 60601-1:2011
- PN-EN 60601-2-52:2010 + AC:2011
- PN-EN 60601-1-6:2010
- PN-EN 62366:2008

This document refers to medical equipment which has been manufactured by Reha- Bed from 02.01.2013 to 31.12.2014

Reha- Bed Management:  
Karolina Staniak: Vice President  
Bartosz Brud: Vice President

  
WICEPREZES ZARZĄDU  
d/s handlu i marketingu  
mgr Karolina Staniak

  
WICEPREZES ZARZĄDU  
mgr inż. Bartosz Brud

<sup>1</sup> COUNCIL DIRECTIVE 93/42/EWG of 14 June 1993 concerning medical devices